

DTU






**DTU Health Technology
Bioinformatics**

Barcoding

*Gisle Vestergaard
Associate Professor
Section of Bioinformatics
Technical University of Denmark
gisves@dtu.dk*

Multiplex sequencing

- Multiplex sequencing allows sequencing libraries to be pooled and sequenced simultaneously during a single run on most high-throughput instrument



- Two representative DNA fragments from two unique samples, each attached to a specific barcode sequence that identifies the sample from which it originated.
- Libraries for each sample are pooled and sequenced in parallel. Each new read contains both the fragment sequence and its sample-identifying barcode.
- Barcode sequences are used to de-multiplex, or differentiate reads from each sample.
- Each set of reads is aligned to the reference sequence.



- NovaSeq system outputs 20 billion paired-end reads yielding up to 3000 Gbases
- Running several samples on one sequencing run
 - Saves money & time
 - Eliminates sequencing run as variable

De-multiplexing

- De-multiplexing is generally done automatically before you receive the samples
- If sequencing locally always check how much on the run ended up unassigned
 - Bcl2fastq2 stringency can be adjusted if barcodes are very different



- How would you realise that demultiplexing was not done?

Index hopping and bleeding

- Failure to remove free adapters leads to index hopping
- Very similar barcodes and sequencing errors can lead to index bleeding
- Dual indexing solves both problems at a cost

